GD-3

APPAREIL DE CONTROLE DE L'ABRASION DES SEMELLES ELASTOMERS ABRASION TESTER

Cet appareil utilisé pour déterminer la résistance à l'abrasion des matériaux élastomères (semelles de chaussures, etc.).
Diamètre du cylindre: 150 mm.
Longueur du cylindre: 500 mm
+ - 40 cycles/ min.
Diamètre de l'éprouvette de 16 mm.
Il possède un écran électronique numérique. Il s'arrête avec une sensibilité de + - 0,2 mètre de la valeur d'essai. L'essai est réalisé entre 1 mètre et 40 +- 0,2 metres. Il est réglable sans échelons. La distance de l'essai unique est de 40 mètres.
La taille du papier de verre utilisé est 400x500 mm.

Thıs device used to determine the abrasion resistance structures of elastomer material (shows sole, etc).
Cylinder diameter: 150 mm.
Cylinder length: 500 mm
+- 40 cycles , cycles / min.
Test specimens diameter 16 mm.
It has electronic digital screen. by pushing the set button within the required tests can stop and the value accuracy the experiment within + - 0.2 meter which at a time of one more tests required standards distance is 40 meters. İt is stepless set.
Used sandpaper size is 400x500 mm

NORMES & STANDARDS
DIN 53516 / DIN EN ISO 20344 / EN 12770

---

info@gateslab.com